Worksheet 5: Quality Assurance

1.	A(n) can be used to check the performance of your instrument and should have a re-				
	covery of between and percent as calculated by the formula				
	·				
2.	A(n) can be used to check variation in the composition of the sample and the				
is a good measure of its difference from the sample, as calculated by the					
	percent.				
2	A(n) can be used to check the effect of adding a known amount of analyte to a				
ე.	A(II) can be used to check the elect of adding a known amount of analyte to a				
	sample solution and should have a recovery of between and percent, as				
	calculated by the formula				

4. Using the table below, calculate:

Run	Concentration (ng/mL)	Absorbance
Blank	0.0	0.0000
Std 1	2.0	0.0175
Std 2	10.0	0.0875
Std 3	20.0	0.1750
Std 4	30.0	0.2625
Std 5	50.0	0.4375
Std 6	75.0	0.6563
Std 7	100.0	0.8750
IPC	50.0	0.4411
Calibration Blank		0.0008
Sample		0.5425
Duplicate Sample		0.5372
LFSM	+10.0	0.6185
IPC	50.0	0.4367
Calibration Blank		0.0004

- (a) The extinction coefficient of EDTA
- (b) The missing concentrations
- (c) The percent recovery of the instrument performance checks

(d)) The	percent	spike	recovery
-----	-------	---------	-------	----------

- (e) The percent relative deviation of the replicate samples
- 5. Comment on the acceptability of the recoveries and deviation calculated above.